

## DESCRIPTION

### Detailed description

[Para 1] The apparatus consists of the stand for holding the binocular and stand for holding the test tubes. The process is consisting in visual analysis of gravitationally separated fractions of particles continuously, in sample of varying granulometrical consistency. Below are the detail description of the apparatus and process:

[Para 2] 1. The vertical to horizontal change in focusing of binocular microscope is achieved by device on FIG. 2. This device is consisting of rotating in horizontal plain bushing 29, with horizontal rotating bar 26, both holding on vertical sliding bushing 25. This device is easily attachable to conventional binocular microscope and without space or time sacrificing completes the conversion.

[Para 3] 2. Apparatus on FIG. 1 achieves the vertical holding of test tube. The base of the device is vertical travel bar 113. The test tube holding plate 114 with the vertical guiding hole 115. The test tube is fixed by rubber ring 120, which is placed in holding groove 121 with slight tension.

[Para 4] 3. The vertical positioning of the test tube plate 114 is achieved by a device, consisting of a vertical screw 111 moving the sliding vertical bushing 112 attached to the plate 114. The rotation is set by vertical movement handle 110.

[Para 5] 4. The process consists of horizontally focusing on the visual sample and analyzing it in liquid. The difference from conventional process of sample analyzing is that in vertical position the gravitationally separated fractions of the sample will be not disturbed during the analysis and study of the sample.

## CROSS REFERENCE TO RELATED APPLICATIONS

[Para 6] References Cited: US Patent application 10/710840 Kosta Zamfes August 05, 2004.

## Field of invention

[Para 7] The invention is related to apparatus and process for visual analyses of gravitationally separated fractions of granular sample in liquid. This apparatus and process are necessary for micro-granulometry analyses that is main part of field and well site analyses of unconsolidated fractions of the formations in sub-strata. Obtaining the quantitative and qualitative degree of fractional separation of particles is beneficial for the physical and petrophysical analyses and formation log interpretation of the formation drilled for Oil and Gas or other targets.

## Background of invention

[Para 8] During drilling of a well the knowledge of micro-granulometrical properties of the unconsolidated formations is one of the primary information in Oil and Gas exploration. The gravitationally separated fractions are easily disturbed if the vertical position of sample is changed to horizontal. For proper Microgranulometry analysis it is necessary to position the sample in the tube and hold it firmly with capability of vertical movement, in front of the Horizontally Focusable Binocular Microscope. Conventionally some of the cutting analyses are done under the vertically focused binocular microscope with great difficulty and low quality. Especially difficult are to analyze the sands and silts within clay matrices and sandy and silty clays/shales. We disclose the ways to obtain the visual analysis of gravitationally separated fractions continuously in sample of varying granulometrical consistency.

## Summary of invention

[Para 9] Apparatus and process of this invention are provided for obtaining the horizontal focusing on binocular microscope and vertical holding of the gravitationally separated sample in the tube for visual analyses of separated fractions of particles without disturbing the sequences.

[Para 10] 1. The vertical to horizontal change in focusing of binocular microscope is achieved by device on FIG. 2. This device is consisting of rotating in horizontal plain bushing 29, with horizontally rotating bar 26, both holding on vertical sliding bushing 25.

[Para 11] 2. Apparatus on FIG. 1 achieves the vertical holding of test tube. The base of the device is vertical travel bar 113. The test tube holding plate 114 with the vertical guiding hole 115. The test tube is fixed by rubber ring 120, which is placed in holding groove 121 with slight tension.

[Para 12] 3. The device consisting of vertical screw 111, which is moving the sliding vertical bushing 112 attached to the plate 114, achieves the vertical positioning of the test tube plate 114. The rotation of screw 111 is set by vertical movement handle 110.

## BRIEF DESCRIPTION OF DRAWINGS

[Para 13] FIG. 1. is a schematic of horizontally focused binocular for vertically gravitated and floating particles samples.

[Para 14] FIG. 2. is a schematic of Assembly for changing the vertical focusing to horizontal.

[Para 15] FIG. 3. is a schematic of Vertical test tube holding.

[Para 16] FIG. 4. is a schematic of Elastic side holder for vertically placed test tubes.